



## CERTIFICATE OF ANALYSIS

---

### TRUETROL ADULTERATION CONTROL LEVEL 4 (NEW FORMULATION)

CATALOG NUMBER: A-1324N, A-1374N      LOT NUMBER: 022615  
EXP. DATE: 02-2016

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 4

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY      Creatinine      98.0%  
Sodium Nitrite      99.4%  
Pyridinium Chlorochromate      99.4%

COLORMETRIC ANALYSIS	UNITS	TARGET	SPECIFICATIONS	RESULTS
----------------------	-------	--------	----------------	---------

USING HITACHI p-Modular:

pH		2.7	2.4 - 3.0	2.8
Specific Gravity		1.023	1.022 - 1.024	1.023
Creatinine	(mg/dL)	1.25	1.12 - 1.38	1.21
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	0	0	0

Kathy A. Chambers

Quality Control

2/26/2015

Date