CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 5

CATALOG NUMBER: A-1325 LOT NUMBER: 102212

EXP. DATE: 10-2013

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 5

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY		
	Creatinine	98.0%
	Sodium Nitrite	98.0%
	Pyridinium Chlorochromate	99.4%

COLORMETRIC ANALYSIS USING HITACHI p-Modular:	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		6	5.10 - 6.90	6.24
Specific Gravity		1.010	1.00850 - 1.01150	1.0103
Creatinine	(mg/dL)	8	6.80 - 9.20	8.287
Nitrite	(mg/L)	185	157.25 - 212.75	184.10
Chromium	(Cr [VI], µg/mL)	0	0	0

Charles A. Hay
Quality Control

10/22/2012
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.