CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 5

CATALOG NUMBER: A-1325

LOT NUMBER: 121514

EXP. DATE: 12-2015

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 5

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY		
	Creatinine	98.0%
	Sodium Nitrite	98.0%
	Pyridinium Chlorochromate	99.4%

COLORMETRIC ANALYSIS USING HITACHI p-Modular:	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		6	5.10 - 6.90	6.10
Specific Gravity		1.010	1.008 - 1.012	1.0100
Creatinine	(mg/dL)	8	6.80 - 9.20	8.23
Nitrite	(mg/L)	185	157.25 - 212.75	191.00
Chromium	(Cr [VI], µg/mL)	0	0	0

Kathy A. Chambers
Quality Control

12/15/2014
Date

Excel.COA1325.121514

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.