CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 6

CATALOG NUMBER: A-1326 & A-1376

LOT NUMBER: 062617

EXP. DATE: 07-2018

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 6

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY

Creatinine	98.0%
Sodium Nitrite	99.4%
Pyridinium Chlorochromate	99.4%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ANALYTE	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		11.6	11.0 - 12.2	11.7
Specific Gravity		1.001	1.000 - 1.003	1.001
Creatinine	(mg/dL)	23	20.7 - 25.3	24.9
Nitrite	(mg/L)	110	99 - 121	111.5
Chromium (Cr [VI])	(µg/mL)	0	0	0
Oxidant	(mg/L)	91	77 - 105	86.9

Kathy A. Chambers 6/26/2017

Quality Control

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.