



## CERTIFICATE OF ANALYSIS

### IDENTIFICATION CONTROL LEVEL 6

CATALOG NUMBER: A-1326 LOT NUMBER: 110413  
EXP. DATE: 11-2014

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 6

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY            Creatinine                            98.0%  
                                 Sodium Nitrite                        99.4%  
                                 Pyridinium Chlorochromate        99.4%

### COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ANALYTE	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		11.6	9.86 - 13.34	11.7
Specific Gravity		1.001	1.000 - 1.003	1.0010
Creatinine	(mg/dL)	23	19.55 - 26.45	24.46
Nitrite	(mg/L)	110	93.5 - 126.5	113.90
Chromium (Cr [VI])	(µg/mL)	0	0	0

Charles A. Hay  
Quality Control

11/4/2013  
Date