



## CERTIFICATE OF ANALYSIS

### TRUETROL ADULTERATION CONTROL LEVEL 6 (NEW FORMULATION)

CATALOG NUMBER: A-1326N LOT NUMBER: 030215  
EXP. DATE: 09-2015

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 6

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%  
Sodium Nitrite 99.4%  
Pyridinium Chlorochromate 99.4%

#### COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ANALYTE	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		11.6	11.0 - 12.2	11.73
Specific Gravity		1.001	1.000 - 1.003	1.001
Creatinine	(mg/dL)	23	20.7 - 25.3	24.48
Nitrite	(mg/L)	110	99 - 121	115.80
Chromium (Cr [VI] )	(µg/mL)	0	0	0

*Kathy A. Chambers*

Quality Control

3/2/2015

Date