



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 5 mg/dL

CATALOG NUMBER: C-1252 & C-1272 LOT NUMBER: 010317
EXP. DATE: 07-2018

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 5 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-mod: mg/dL	5	4.5 - 5.5	4.996

Kathy A. Chambers
Quality Control

1/3/2017
Date