



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 5 mg/dL

CATALOG NUMBER: C-1252 & C-1272 LOT NUMBER: 021521
EXP. DATE: 08-31-2022

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 5 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 100.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-mod: mg/dL	5	4.25 - 5.75	5.00

Terra S. Brown

Quality Control

2/16/2021

Date
