



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 2 mg/dL

CATALOG NUMBER: C-1254 LOT NUMBER: 011316

EXT. EXP. DATE: 01-2017

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 2 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/dL	2	1.8 - 2.2	2.057

Kathy A. Chambers

Quality Control

1/13/2016

Date