



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 2 mg/dL

CATALOG NUMBER: C-1254 & C-1274 LOT NUMBER: 030520

EXT. EXP. DATE: 03-31-21

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 2 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/dL	2	1.7 - 2.3	1.98

Kathy A. Chambers

Quality Control

3/5/2020

Date