



## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ CREATININE CALIBRATOR 2 mg/dL

CATALOG NUMBER: C-1254 LOT NUMBER: 041315  
EXT. EXP. DATE: 04-2016

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 2 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

#### COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/dL	2	1.8 - 2.2	2.058

Kathy A. Chambers

Quality Control

4/13/2015

Date