



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 2 mg/dL

CATALOG NUMBER: C-1254 LOT NUMBER: 102113
EXT. EXP. DATE: 10-2014

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 2 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/dL	2	1.7 - 2.3	2.026

Charles A. Hay
Quality Control

10/21/2013
Date