CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 2 mg/dL

CATALOG NUMBER: C-1254 & C-1274 LOT NUMBER: 20030217

EXT. EXP. DATE: 03-2018

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 2 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/dL	2	1.8 - 2.2	2.058

Kathy A. Chambers 3/2/2017

Quality Control Date

Excel.COA1254.XLS.030217

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.