CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 2 mg/dL

CATALOG NUMBER: C-1254 & C-1274 LOT NUMBER: 20040518

EXT. EXP. DATE: 04-2019

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 2 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/dL	2	1.8 - 2.2	2.041

Kathy A. Chambers 4/5/2018

Quality Control Date

Excel.COA1254.XLS.20040518

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.