



## CERTIFICATE OF ANALYSIS

### TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL

CATALOG NUMBER: C-1257 & C-1277 LOT NUMBER: 032219

EXP. DATE: 04-2020

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:			
mg/dL	200	180 - 220	198.5

*Kathy A. Chambers*

Quality Control

4/2/2019

Date