CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL

CATALOG NUMBER: C-1257 & C-1277

LOT NUMBER: 041316

EXP. DATE: 04-2017

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS USING HITACHI p-Modular:		TARGET	SPECIFICATIONS	RESULTS
	mg/dL	200	180 - 220	204.33

Kathy A. Chambers
Quality Control

4/13/2016
Date

Excel.COA1257.041316

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.