



## CERTIFICATE OF ANALYSIS

### TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL

CATALOG NUMBER: C-1257 & C-1277

LOT NUMBER: 042020

EXP. DATE: 05-31-2021

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:			
mg/dL	200	170 - 230	200

Kathy A. Chambers

Quality Control

4/20/2020

Date