



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL

CATALOG NUMBER: C-1257 LOT NUMBER: 051214
EXP. DATE: 05-2015

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI 717:			
mg/dL	200	170 - 230	189.95

Charles A. Hay

Quality Control

5/12/2014

Date