



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL

CATALOG NUMBER: C-1257 & C-1277

LOT NUMBER: 051721

EXP. DATE: 05-31-2022

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 100.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:			
mg/dL	200	170 - 230	198

Terra S. Brown

Quality Control

5/18/2021

Date