



**CERTIFICATE OF ANALYSIS**

**TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL**

CATALOG NUMBER: C-1257 & C-1277

LOT NUMBER: 063015

EXP. DATE: 06-2016

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI 717:			
mg/dL	200	180 - 220	202.97

Kathy A. Chambers

Quality Control

6/30/2015

Date