



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 200 mg/dL

CATALOG NUMBER: C-1257 & C-1277

LOT NUMBER: 100319

EXP. DATE: 10-31-2020

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 200 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:			
mg/dL	200	180 - 220	198

Kathy A. Chambers

Quality Control

10/3/2019

Date