CERTIFICATE OF ANALYSIS

TEST TRUE™ CHROMATE CALIBRATOR 50 mg/L

CATALOG NUMBER: M-1151 & M-1171 LOT NUMBER: 062420

EXP. DATE: 07-31-2021

PRODUCT NAME: TEST TRUE CHROMATE CALIBRATOR 50 mg/L

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY Pyridinium chlorochromate 98.0%

COLORMETRIC ANALYSIS		TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:	mg/L	50	42.5 - 57.5	50.3

Terra S. Brown 6/30/2020

Quality Control Date

Excel.COA1151.XLS.062420

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.