



---

## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ CHROMATE CALIBRATOR 50 mg/L

CATALOG NUMBER: M-1151 & M-1171 LOT NUMBER: 091219  
EXP. DATE: 09-30-2020

PRODUCT NAME: TEST TRUE CHROMATE CALIBRATOR 50 mg/L

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY Pyridinium chlorochromate 98.0%

COLORMETRIC ANALYSIS	TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:	mg/L 50	45 - 55	49.9

Kathy A. Chambers

Quality Control

9/12/2019

Date

---