CERTIFICATE OF ANALYSIS

TEST TRUE™ CHROMATE CALIBRATOR 50 mg/L

CATALOG NUMBER: M-1151 LOT NUMBER: 121012

EXP. DATE: 12-2013

PRODUCT NAME: TEST TRUE CHROMATE CALIBRATOR 50 mg/L

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY Pyridinium chlorochromate 98.0%

COLORMETRIC ANALYSIS		TARGET	SPECIFICATIONS	RESULTS
USING HITACHI p-Modular:	mg/L	50	42.5 - 57.5	49.60

Charles A. Hay
Quality Control

12/17/2012
Date

Excel.COA1151.XLS.121012

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.