



CERTIFICATE OF ANALYSIS

TEST TRUE™ NITRITE CALIBRATOR 500 mg/L

CATALOG NUMBER: N-1011 & N-1071 LOT NUMBER: 020917
EXT. EXP. DATE: 02-2018

PRODUCT NAME: TEST TRUE NITRITE CALIBRATOR 500 mg/L

ANALYTE PURITY Sodium Nitrite 98.0%

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular :

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	500	450 - 550	500.8

Kathy A. Chambers

Quality Control

2/9/2017

Date
