



---

## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ NITRITE CALIBRATOR 500 mg/L

CATALOG NUMBER: N-1011 & N-1071                      LOT NUMBER: 072120  
EXT. EXP. DATE: 08-31-2021

PRODUCT NAME: TEST TRUE NITRITE CALIBRATOR 500 mg/L

ANALYTE PURITY                      Sodium Nitrite                      98.0%

APPEARANCE:                      CLEAR, COLORLESS SOLUTION

#### COLORMETRIC ANALYSIS

USING HITACHI p-Modular :

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	500	425 - 575	514

Terra S. Brown

Quality Control

7/31/2020

Date

---