



CERTIFICATE OF ANALYSIS

TEST TRUE™ NITRITE CALIBRATOR 500 mg/L

CATALOG NUMBER: N-1011 & N-1071 LOT NUMBER: 20020320
EXT. EXP. DATE: 01-31-2021

PRODUCT NAME: TEST TRUE NITRITE CALIBRATOR 500 mg/L

ANALYTE PURITY Sodium Nitrite 98.0%

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular :

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	500	450 - 550	504

Kathy A. Chambers

Quality Control

2/5/2020

Date
