



CERTIFICATE OF ANALYSIS

TEST TRUE™ NITRITE CALIBRATOR 500 mg/L

CATALOG NUMBER: N-1011 & N-1071 LOT NUMBER: 20080317
EXT. EXP. DATE: 08-2018

PRODUCT NAME: TEST TRUE NITRITE CALIBRATOR 500 mg/L

ANALYTE PURITY Sodium Nitrite 98.0%

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular :

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	500	450 - 550	508.3

Kathy A. Chambers

Quality Control

8/3/2017

Date
