



## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ OXIDANT (NITRITE) CALIBRATOR 150 mg/L

CATALOG NUMBER: O-1352 & O-1372                      LOT NUMBER: 050722  
EXT. EXP. DATE: 11-30-2023

PRODUCT NAME: TEST TRUE OXIDANT (NITRITE) CALIBRATOR 150 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY                      Sodium Nitrite                      98.7%

#### COLORMETRIC ANALYSIS

USING HITACHI p Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	150	127.5 - 172.5	153.6

Terra S. Brown

Quality Control

5/19/2022

Date

---