



## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ OXIDANT (NITRITE) CALIBRATOR 150 mg/L

CATALOG NUMBER: O-1352 & O-1372                      LOT NUMBER: 062217  
EXT. EXP. DATE: 12-2018

PRODUCT NAME: TEST TRUE OXIDANT (NITRITE) CALIBRATOR 150 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY                      Sodium Nitrite                      98.0%

#### COLORMETRIC ANALYSIS

USING HITACHI p Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	150	135 - 165	150.8

Kathy A. Chambers

Quality Control

6/22/2017

Date

---