



CERTIFICATE OF ANALYSIS

TEST TRUE™ OXIDANT (NITRITE) CALIBRATOR 150 mg/L

CATALOG NUMBER: O-1352 LOT NUMBER: 100412
EXT. EXP. DATE: 04-2014

PRODUCT NAME: TEST TRUE OXIDANT (NITRITE) CALIBRATOR 150 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Sodium Nitrite 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p Modular:

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	150	127.50 - 172.50	150.18

Charles A. Hay

Quality Control

10/4/2012

Date