# CERTIFICATE OF ANALYSIS

**TEST TRUE™ OXIDANT (NITRITE) CALIBRATOR 150 mg/L**

CATALOG NUMBER: O-1352 LOT NUMBER: 111213

EXT. EXP. DATE: 05-2015

PRODUCT NAME: TEST TRUE OXIDANT (NITRITE) CALIBRATOR 150 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Sodium Nitrite 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p Modular:

| UNITS | TARGET | SPECIFICATIONS | RESULTS |
|---|---|---|---|
| mg/L | 150 | 127.50 - 172.50 | 149.51 |

*Charles A. Hay*
Quality Control

11/12/2013
Date



Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.