



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1051N & P-1071N LOT NUMBER: 040521
EXP. DATE: 04-30-2022

PRODUCT NAME: TEST TRUE pH 3.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	3.0	2.7 - 3.3	2.96

Terra S. Brown

Quality Control

4/7/2021

Date