



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1051N P-1071N LOT NUMBER: 042815
EXP. DATE: 05-2016

PRODUCT NAME: TEST TRUE pH 3.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	3.0	2.7 - 3.3	3.01

Kathy A. Chambers

Quality Control

4/28/2015

Date
