



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1051N & P-1071N LOT NUMBER: 091321
EXP. DATE: 09-30-2022

PRODUCT NAME: TEST TRUE pH 3.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	3.0	2.7 - 3.3	3.00

Terra S. Brown

Quality Control

9/17/2021

Date