



## CERTIFICATE OF ANALYSIS

---

CATALOG NUMBER: P-1052N & P-1072N      LOT NUMBER: 022117  
EXP. DATE: 03-2018

PRODUCT NAME: TEST TRUE pH 11.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS  
USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	11.0	10.7 - 11.3	10.99

Kathy A. Chambers

Quality Control

2/21/2017

Date