



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1052N P-1072N LOT NUMBER: 042815

EXP. DATE: 05-2016

PRODUCT NAME: TEST TRUE pH 11.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	11.0	10.7 - 11.3	11.010

Kathy A. Chambers

Quality Control

4/28/2015

Date
