



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1052N & P-1072N LOT NUMBER: 091917

EXP. DATE: 10-2018

PRODUCT NAME: TEST TRUE pH 11.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	11.0	10.7 - 11.3	11.10

Kathy A. Chambers

Quality Control

9/19/2017

Date
