



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1052N & P-1072N LOT NUMBER: 112119
EXP. DATE: 11-30-2020

PRODUCT NAME: TEST TRUE pH 11.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS
USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	11.0	10.7 - 11.3	10.99

Kathy A. Chambers

Quality Control

11/25/2019

Date