



CERTIFICATE OF ANALYSIS

CATALOG NUMBER: P-1052N & P-1072N LOT NUMBER: 20100420

EXP. DATE: 10-31-2021

PRODUCT NAME: TEST TRUE pH 11.0 CALIBRATOR

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

	TARGET	SPECIFICATIONS	RESULTS
pH	11.0	9.9 - 12.1	10.99

Terra S. Brown

Quality Control

10/9/2020

Date