CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 1

CATALOG NUMBER: T-3001 & T-3017

LOT NUMBER: 011121

EXP. DATE: 01-31-2022

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 1

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY

Creatinine	100.0%
Pyridinium Chlorochromate	99.4%

COLORMETRIC ANALYSIS

USING HITACHI p-modular:

	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH	(mg/L)	3.6	3.1 - 4.1	3.7
Specific Gravity	(mg/L)	1.023	1.019 - 1.027	1.023
Creatinine	(mg/dL)	1.25	1.06 - 1.44	1.24
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	62	53 - 71	61.1

Terra S. Brown
Quality Control

1/14/2021
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.