# CERTIFICATE OF ANALYSIS

**TRUETROL ADULTERATION CONTROL LEVEL 1**

CATALOG NUMBER: T-3001 & T-3017

LOT NUMBER: 031319

EXP. DATE: 03-2020

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 1

APPEARANCE: CLEAR, YELLOW SOLUTION

| ANALYTE PURITY | | |
|---|---|---|
| | Creatinine | 98.0% |
| | Pyridinium Chlorochromate | 99.4% |

COLORMETRIC ANALYSIS

USING HITACHI p-modular:

| | UNITS | TARGET | SPECIFICATIONS | RESULTS |
|---|---|---|---|---|
| pH | (mg/L) | 3.6 | 3.2 - 4.0 | 3.6 |
| Specific Gravity | (mg/L) | 1.023 | 1.019 - 1.027 | 1.023 |
| Creatinine | (mg/dL) | 1.25 | 1.12 - 1.38 | 1.29 |
| Nitrite | (mg/L) | 0 | 0 | 0 |
| Chromium | (Cr [VI], µg/mL) | 62 | 56 - 68 | 62.1 |

*Kathy A. Chambers*
Quality Control

3/13/2019
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.