CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 1

CATALOG NUMBER: T-3001 & T-3017

LOT NUMBER: 033020

EXP. DATE: 04-30-2021

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 1

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY		
	Creatinine	98.0%
	Pyridinium Chlorochromate	99.4%

COLORMETRIC ANALYSIS

USING HITACHI p-modular:

	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH	(mg/L)	3.6	3.1 - 4.1	3.6
Specific Gravity	(mg/L)	1.023	1.019 - 1.027	1.023
Creatinine	(mg/dL)	1.25	1.06 - 1.44	1.29
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	62	53 - 71	64.5

Kathy A. Chambers
Quality Control

4/6/2021
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.