CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 1

CATALOG NUMBER: T-3001 & T-3017

LOT NUMBER: 062321

EXP. DATE: 06-30-2022

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 1

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY

Creatinine	100.0%
Pyridinium Chlorochromate	99.4%

COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON HITACHI P-MODULAR ANALYZER. THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIOM RECOMMENDS THAT YOUR FACILITY ESTABLISH IT'S OWN RANGE ON EACH ANALYZER.

	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH	(mg/L)	3.6	3.1 - 4.1	3.6
Specific Gravity	(mg/L)	1.023	1.019 - 1.027	1.023
Creatinine	(mg/dL)	1.25	1.06 - 1.44	1.32
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	62	53 - 71	66

Terra S. Brown
Quality Control

7/6/2021
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.