CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 1

CATALOG NUMBER: T-3001 & T-3017 LOT NUMBER: 21071130

EXP. DATE: 07-31-2022

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 1

APPEARANCE: CLEAR, YELLOW SOLUTION

ANALYTE PURITY Creatinine 100.0%

Pyridinium Chlorochromate 99.4%

COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON HITACHI P-MODULAR ANALYZER. THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIM RECOMMENDS THAT YOUR FACILITY ESTABLISH IT'S OWN RANGE ON EACH ANALYZER.

	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH	(mg/L)	3.6	3.1 - 4.1	3.6
Specific Gravity	(mg/L)	1.023	1.019 - 1.027	1.023
Creatinine	(mg/dL)	1.25	1.06 - 1.44	1.20
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	62	53 - 71	64.3

Terra S. Brown 9/10/2021

Quality Control Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.