



CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL T LEVEL 2

CATALOG NUMBER: T-3002 & T-3027 LOT NUMBER: 020922
EXP. DATE: 02-28-2023
PRODUCT NAME: TEST TRUE TRUETROL CONTROL T LEVEL 2
APPEARANCE: CLEAR, COLORLESS SOLUTION
ANALYTE PURITY CREATININE 100.0%

COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON HITACHI P-MODULAR ANALYZER.
THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIOM RECOMMENDS THAT
YOUR FACILITY ESTABLISH IT'S OWN RANGE ON EACH ANALYZER.

<u>ITEM</u>	<u>UNITS</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
pH		4.25	3.61 - 4.89	4.40
Specific Gravity		1.017	1.014 - 1.020	1.017
Creatinine	(mg/dL)	3.5	2.98 - 4.03	3.32
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	0	0	0

Terra S. Brown

Quality Control

2/9/2022

Date



Axiom is an ISO 13485 Certified Company for Medical Devices by NQA

Axiom's controls and calibrators are manufactured and certified under a QMS that is accredited to ISO 13485. Calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.