# CERTIFICATE OF ANALYSIS

**TRUETROL ADULTERATION CONTROL LEVEL 2**

CATALOG NUMBER: T-3002 LOT NUMBER: 030314

EXP. DATE: 09-2015

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 2

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

| ITEM | UNITS | TARGET | SPECIFICATIONS | RESULTS |
|---|---|---|---|---|
| pH | | 3.60 | 3.06 - 4.14 | 3.60 |
| Specific Gravity | | 1.01700 | 1.01445 - 1.01955 | 1.01700 |
| Creatinine | (mg/dL) | 3.5 | 2.98 - 4.03 | 3.59 |
| Nitrite | (mg/L) | 0 | 0 | 0 |
| Chromium | (Cr [VI], µg/mL) | 0 | 0 | 0 |

*Charles A. Hay* 3/3/2014

Quality Control Date



Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.