



## CERTIFICATE OF ANALYSIS

### TRUETROL ADULTERATION CONTROL LEVEL 2

CATALOG NUMBER: T-3002 LOT NUMBER: 041713  
EXP. DATE: 10-2014

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 2

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

#### COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		3.60	3.06 - 4.14	3.60
Specific Gravity		1.01700	1.01445 - 1.01955	1.01690
Creatinine	(mg/dL)	3.5	2.98 - 4.03	3.50
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	0	0	0

*Charles A. Hay*

Quality Control

*4/17/2013*

Date