



CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 2

CATALOG NUMBER: T-3002 & T-3027 LOT NUMBER: 082118
EXP. DATE: 09-2019

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 2

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		4.25	3.85 - 4.65	4.20
Specific Gravity		1.017	1.014 - 1.020	1.017
Creatinine	(mg/dL)	3.5	3.15 - 3.85	3.57
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	0	0	0

Kathy A. Chambers

Quality Control

8/21/2018

Date