



CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 2

CATALOG NUMBER: T-3002 & T-3027 LOT NUMBER: 101519
EXP. DATE: 10-31-2020

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 2

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		4.25	3.61 - 4.89	4.34
Specific Gravity		1.017	1.014 - 1.020	1.017
Creatinine	(mg/dL)	3.5	2.98 - 4.03	3.56
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	0	0	0

Kathy A. Chambers

Quality Control

10/15/2019

Date